



LL 190-201
MW = 1383.5
Catalog #: PI-L-04
Lot #: K0807009

Mode of operation: Linear
Extraction mode: Delayed
Polarity: Negative
Acquisition control: Manual
Accelerating voltage: 20000 V
Grid voltage: 90%
Guide wire 0: 0.05%
Extraction delay time: 100 nsec
Acquisition mass range: 850 -- 5000 Da
Number of laser shots: 300/spectrum
Laser intensity: 677
Laser Rep Rate: 20.0 Hz
Calibration type: Default
Calibration matrix: a-Cyano-4-hydroxycinnamic acid
Low mass gate: 500 Da
Digitizer start time: 18.46
Bin size: 2 nsec
Number of data points: 13077
Vertical scale: 200 mV
Vertical offset: -5%
Input bandwidth: 150 MHz
Sample well: 04
Plate ID: 100 WELL
Serial number: 1215
Instrument name: Voyager-DE
Plate type filename: C:\VOYAGER\100 well plate.pt
Lab name: PE Biosystems
Absolute x-position: 16242.3
Absolute y-position: 47655.9
Relative x-position: -585.172
Relative y-position: 348.355
Shots in spectrum: 80
Source pressure: 5.036e-007
Mirror pressure: 0
TC2 pressure: 0.001
TIS gate width: 30
TIS flight length: 940