



Mode of operation: Linear
 Extraction mode: Delayed
 Polarity: Negative
 Acquisition control: Manual

Accelerating voltage: 20000 V
 Grid voltage: 90%
 Guide wire 0: 0.05%
 Extraction delay time: 100 nsec

Acquisition mass range: 800 -- 6000 Da
 Number of laser shots: 300/spectrum
 Laser intensity: 794
 Laser Rep Rate: 20.0 Hz
 Calibration type: Default
 Calibration matrix: a-Cyano-4-hydroxycinnamic acid
 Low mass gate: 750 Da

Digitizer start time: 17.912
 Bin size: 2 nsec
 Number of data points: 15474
 Vertical scale: 200 mV
 Vertical offset: -5%
 Input bandwidth: 150 MHz

Sample well: 23
 Plate ID: 100 WELL
 Serial number: 1215
 Instrument name: Voyager-DE
 Plate type filename: C:\VOYAGER\100 well plate.pt
 Lab name: PE Biosystems

Absolute x-position: 13284.5
 Absolute y-position: 37415.2
 Relative x-position: 1537.04
 Relative y-position: 267.702
 Shots in spectrum: 71
 Source pressure: 1.293e-006
 Mirror pressure: 0
 TC2 pressure: 0.00145
 TIS gate width: 30
 TIS flight length: 940