

Voyager Spec #1[BP = 2517.1, 35150]

Mode of operation: Linear
Extraction mode: Delayed
Polarity: Negative
Acquisition control: Manual

Accelerating voltage: 20000 V
Grid voltage: 94%
Guide wire 0: 0.07%
Extraction delay time: 200 nsec

Acquisition mass range: 800 -- 5000 Da
Number of laser shots: 100/spectrum
Laser intensity: 2212
Laser Rep Rate: 20.0 Hz
Calibration type: External -- C:\VOYAGER\Data\ACTH(18-39)_cal.
Calibration matrix: a-Cyano-4-hydroxycinnamic acid
Low mass gate: 500 Da

Digitizer start time: 17.99
Bin size: 2 nsec
Number of data points: 13401
Vertical scale: 1000 mV
Vertical offset: 0%
Input bandwidth: 150 MHz

Sample well: 91
Plate ID: PLATE 3
Serial number: 1215
Instrument name: Voyager-DE
Plate type filename: C:\VOYAGER\100 well plate.plt
Lab name: PE Biosystems

Absolute x-position: 2021.18
Absolute y-position: 1096.44
Relative x-position: 433.68
Relative y-position: -491.063
Shots in spectrum: 100
Source pressure: 1.512e-007
Mirror pressure: 0
TC2 pressure: 0.004171
TIS gate width: 30
TIS flight length: 940

